

## ThawSTAR® CB **Quick Start Guide**































Wait for thaw to complete

▲ Stay clear of moving drawer




Select "Finish" to close drawer



### ThawSTAR® CB Notice: Emergency Procedure

#### Follow steps for bag removal in emergency:



# Restoring power after emergency procedure

- 1 Return emergency release lever back to its starting position
- 2 Replace emergency lever cover
- 3 Reconnect power cord and turn power switch on
- 4 Leave drawer open it will close automatically on power up



3303 Monte Villa Parkway Suite 310 Bothell, WA 98021 USA

- T +1.866.424.6543 / +1.425.402.1400
- F +1.425.402.1433
- E ThawTechSupport@BioLifeSolutions.com

BioLifeSolutions.com

2025@ BioLife Solutions.
All rights reserved.

#### Water-free Thawing

QG-AST-250897\_R03 CFD-0038\_1.0